Virta Classic

PRODUCT SPECIFICATION

PRODUCT DESCRIPTION					
Brand Name	**Virta Classic** Nitrile Powder Free, Non Sterile Examination Gloves.				
Intentded Use	A patient examination gloved is a disposable device intended for medical purpose that is worn on the examiner's hand or finger to prevent contamination between patient and examiner				
Product Conformance	ASTM D6319 current revision as per Customer requirement				
Material	Nitrile (Carboxylated Acrylonitrile Butadiene)				
Cuff Finishing	Beaded				
Color	Blue				
Type	Ambidexterous				
Surface Finish	Fingertip Textured				
SPECIFICATIONS					
Dimension	**Size**	**Palm Width**	**Length (mm)**		
	S	86 ± 3	≥240		
	M	98 ± 3	≥240		
	L	107 ± 3	≥240		
	XL	115 ± 3	≥240		
Thickness		**Single Wall (mm) Specification**			
	Cuff	0.05min			
	Palm	0.06min			
	Finger	0.07min			
Physical Properties		**Un-Aged Specification**		**Aged Specification**	
	Tensile Strength (MPa)	14min		14min	
	Elongation (%)	500min		400min	
Residual Powder	2.0mg / glove maximum				
Quality Specification *AND – Accept no Defects	**Inspection Criteria**	**Sampling Plan/AQL**			
	1000mL watertightness	GI, AQL 1.5			
	Physical dimensions	S2, AQL 4.0			
	Physical properties	S2, AQL 4.0			
	Visual Defects – Critical	GI, AQL 0.65			
	Visual Defects -Major	GI, AQL 2.5			
	Visual Defects – Minor	GI, AQL 4.0			
PACKAGE CONFIGURATION AND LABELLING					
Packing	100 gloves (by weight) per inner dispenser. 10 dispensers per outer carton				
Labelling	Dispensers and outer cartons shall have the following information: • Product description including size and quantity of gloves • Country of origin of Manufacturer • Single use only symbol, and non-sterile • Lot Number and date of manufacture				

103 – 251 Saulteaux Cres | Winnipeg, MB | R3J 3C7 |
ph. 800.847.3160 | e. qa@virtagloves.com