Virta Knight

PRODUCT SPECIFICATION

PRODUCT DESCRIPTION

Brand Name	**Virta Knight** Nitrile Powder Free, Non Sterile Examination Gloves.
Intentded Use	A patient examination gloved is a disposable device intended for medical purpose that is worn on the examiner's hand or finger to prevent contamination between patient and examiner
Product Conformance	ASTM D6319 current revision as per Customer requirement
Material	Nitrile (Carboxylated Acrylonitrile Butadiene)
Cuff Finishing	Beaded
Color	Black
Type	Ambidexterous
Surface Finish	Fully Textured

SPECIFICATIONS

Dimension	**Size**	**Palm Width**	**Length (mm)**		
	XS	80± 4	240-260		
	S	86 ± 4	240-260		
	M	98 ± 4	240-260		
	L	108 ± 4	240-260		
	XL	115 ± 4	240-260		
	XXL	126 ± 4	240-260		

Thickness		**Single Wall (mm) Specification**	
	Cuff	0.06-0.11	
	Palm	0.10-0.15	
	Finger	0.12-0.19	
	Beading	0.5min	
Physical Properties		**Un-Aged Specification**	**Aged Specification**
	Tensile Strength (MPa)	18min	16min
	Ultimate Elongation (%)	500min	400min

Residual Powder	2.0mg / glove maximum	
Quality Specification	**Inspection Criteria**	**Sampling Plan/AQL**
	Dimension	S2 – AQL 4.0
	Physical Properties	S2 – AQL 4.0
	Water Tight Test	G1 – AQL 1.0 (ASTM) G11 – AQL 1.5 Multiple
	Visual Inspection – Critical Defects	G1 – AQL 1.5
	Visual Inspection – Major	G1 – AQL 2.5
	Visual Inspection – Minor	G1 – AQL 4.0
	Residual Powder	n = 5 (average)
	Packaging Labelling	n = 5 (accept, no defects)
	Glove Physical Counting	n = 5 (accept, no defects)

PACKAGE CONFIGURATION AND LABELLING

Packing	100 gloves (by weight) per inner dispenser. 10 dispensers per outer carton
Labelling	Dispensers and outer cartons shall have the following information: • Product description including size and quantity of gloves • Country of origin of Manufacturer • Single use only symbol, and non-sterile • Lot Number and date of manufacture

103 – 251 Saulteaux Cres | Winnipeg, MB | R3J 3C7 |
ph. 800.847.3160 | e. qa@virtagloves.com