Virta. Soothe

# PRODUCT SPECIFICATION

| PRODUCT DESCRIPTION | | | | | |
|---|---|---|---|---|---|
| **Brand Name** | **Virta Soothe** Nitrile Powder Free, Non Sterile Examination Gloves. | | | | |
| **Intentded Use** | A patient examination gloved is a disposable device intended for medical purpose that is worn on the examiner's hand or finger to prevent contamination between patient and examiner | | | | |
| **Product Conformance** | EN 455, ASTM D6319, ASTM D6978 current revision as per Customer requirement | | | | |
| **Material** | Nitrile (Carboxylated Acrylonitrile Butadiene) | | | | |
| **Cuff Finishing** | Beaded | | | | |
| **Color** | Jade Green | | | | |
| **Type** | Ambidexterous | | | | |
| **Surface Finish** | Fingertip Textured | | | | |
| **SPECIFICATIONS** | | | | | |
| **Dimension** | **Size** | **Palm Width** | **Length (mm)** | | |
| | XS | 75± 5 | 250min | | |
| | S | 85 ± 5 | 250min | | |
| | M | 95 ± 5 | 250min | | |
| | L | 106 ± 5 | 250min | | |
| | XL | 115 ± 5 | 250min | | |
| **Thickness** | | **Single Wall (mm) Specification** | | | |
| | Cuff | 0.05-0.08 | | | |
| | Palm | 0.06-0.09 | | | |
| | Finger | 0.08-0.14 | | | |
| | Beading | 0.5min | | | |
| **Physical Properties** | | **Un-Aged Specification** | **Aged Specification** | | |
| | Force at Break (N) | 6 min | 6 min | | |
| **Residual Powder** | 2.0mg / glove maximum | | | | |
| **Quality Specification**<br><br>*AND – Accept no Defects | **Inspection Criteria** | **Sampling Plan/AQL** | | | |
| | Dimension | 13 pcs (Median) | | | |
| | Force at Break | 13 pcs (Median) | | | |
| | Water Tight Test | AND – 100% Screen | | | |
| | Residual Powder | n = 5 (average) | | | |
| | Packaging Labelling | n = 5 (AND) | | | |
| | Glove Physical Counting | n = 5 (AND) | | | |
| **PACKAGE CONFIGURATION AND LABELLING** | | | | | |
| **Packing** | 200 gloves (by weight) per inner dispenser. 10 dispensers per outer carton | | | | |
| **Labelling** | Dispensers and outer cartons shall have the following information:<br>Product description including size and quantity of gloves<br>Country of origin of Manufacturer<br>Single use only symbol, and non-sterile<br>Lot Number and date of manufacture | | | | |



103 – 251 Saulteaux Cres | Winnipeg, MB | R3J 3C7 |
ph. 800.847.3160 | e. qa@virtagloves.com

Virta. Soothe

# PRODUCT SPECIFICATION

| PRODUCT DESCRIPTION | | | | | |
|---|---|---|---|---|---|
| Brand Name | Virta Soothe Nitrile Powder Free, Non Sterile Examination Gloves. | | | | |
| Intentded Use | A patient examination gloved is a disposable device intended for medical purpose that is worn on the examiner's hand finger to prevent contamination between patient and examiner | | | | |
| Product Conformance | EN 455, ASTM D6319, ASTM D6978current revision as per Customer requirement | | | | |
| Material | Nitrile (Carboxylated Acrylonitrile Butadiene) | | | | |
| Cuff Finishing | Beaded | | | | |
| Color | Jade Green | | | | |
| Type | Ambidexterous | | | | |
| Surface Finish | Fingertip Textured | | | | |
| SPECIFICATIONS | | | | | |
| Dimension | Size | Palm Width | Length (mm) | | |
| | XS | 75 ± 5 | 250min | | |
| | S | 85 ± 5 | 250min | | |
| | M | 95 ± 5 | 250min | | |
| | L | 10 6 ± 5 | 250min | | |
| | XL | 1 15 ± 5 | 250min | | |
| Thickness | | Single Wall (mm) Specification | | | |
| | Cuff | 0.05-0.08 | | | |
| | Palm | 0.06-0.09 | | | |
| | Finger | 0.08-0.1 4 | | | |
| | Beading | 0.5min | | | |
| Physical Properties | | Un-Aged Specification | | Aged Specification | |
| | Force at Break (N) | 6 min | | 6 min | |
| Residual Powder | 2.0mg / glove maximum | | | | |
| Quality Specification<br><br>*AND– Accept no Defects | Inspection Criteria | | Sampling Plan/AQL | | |
| | Dimension | | 13 pcs (Median) | | |
| | Force at Break | | 13 pcs (Median) | | |
| | Water Tight Test | | AND– 100% Screen | | |
| | Residual Powder | | n = 5 (average) | | |
| | Packaging Labelling | | n = 5 (AND) | | |
| | Glove Physical Counting | | n = 5 (AND) | | |
| PACKAGE CONFIGURATION AND LABELLING | | | | | |
| Packing | 200 gloves (by weight) per inner dispenser. 10 dispensers per outer carton | | | | |
| Labelling | Dispensers and outer cartons shall have the following information:<br>Product description including size and quantity of gloves<br>Country of origin of Manufacturer<br>Single use only symbol, and non-sterile<br>Lot Number and date of manufacture | | | | |



103 – 251 Saulteaux Cres | Winnipeg, MB | R3J 3C7 |
ph. 800.847.3160 | e. qa@virtagloves.com